

## HCC

### Dental Questions RFP# HCC-20200722-D

Updated FAQ V2 8/17/2020

**Question #1** Does HCC of Rural Missouri have a preferred floor plan or layout that can be shared?

**Response #1** Please design your own floor plan with COVID guidelines.

**Question #2** Will HCC accept a 40' unit in lieu of a 45' if interior conversion requirements can be met in a lesser length vehicle?

**Response #2** Yes

**Question #3** Will HCC be doing telehealth on via the desired communication package? If so, please specify the telehealth equipment and desired capabilities.

**Response #3** No

**Question #4** How many CAT 6 data outlets shall be included in the bid? What are the preferred locations?

**Response #4** Assuming that they can be utilized for internet connection, one for each of the operatories and enough in the intake and pano areas to handle the pano computer connection and one computer for the other Xray's in the intake room

**Question #5** Will HCC accept a TV larger than 24"?

**Response #5** Yes

**Question #6** What is the required location of the TV?

**Response #6** They would prefer monitors to be mounted in each operatory in patient eye view for connection to the provider laptop. Third screen would be viewing area for pano and Xray room.

**Question #7** What is the desired location/ functionality of the electronic message sign? Interior? Exterior?

**Response #7** Exterior. Communication with those waiting outside.

**Question #8** If a lead-lined (equivalency) room is provided, does HCC still require a lead-lined curtain?

**Response #8** No. Radiological safety standards are primary concern.

**Question #9** Does HCC have a preference make/model for the evacuation system?

**Response #9** No

**Question #10** Does HCC have a preference make/model for the panoramic X-Ray unit and corresponding software? 2D? 3D?

**Response #10** 2D, Compatible with Xray vision/apteryx

**Question #11** Does HCC have a preference make/model for the self-contained mobile dental unit(s)?

**Response #11** Adec preferred but it is not a sticking point.

**Question #12** Does HCC have a preference make/model for the sterilizer? M9? M11? Other?

**Response #12** M11 please.

**Question #13** Does HCC have a preference make/model for the ultrasonic cleaner? 1.2g? 3g? Other?

**Response #13** No

**Question #14** Does HCC have a preference make/model for the X-Ray Apron(s)?

**Response #14** No

**Question #14** Shall hooks or other storage accommodations be made for the X-Ray apron(s)?

**Response #14** Yes

**Question #15** What size sensors does HCC desire for the Hand Held X-Ray Unit? Size O? Size 1? Other?

**Response #15** One size 1 and one size 2. Gendex preferred. Must be compatible with apteryx/Xray vision software.

**Question #16** Page 13 of 14 - Will a 33ft 50amp powered shore reel with 25' 50amp extension cord be acceptable?

**Response #16** Yes