

HCC

Medical RFP# HCC-20200722-M

Updated FAQ V2 8/17/2020

Question #1 Does HCC of Rural Missouri have a preferred floor plan or layout that can be shared?

Response #1 Please design your own floor plan with COVID guidelines.

Question #2 Does HCC have a preference for inclusions in the exam room:

Response #2 Exam Table Make/Model preference? No preference.

Hi-Low, Zero Barrier? No preference.

Box Type Table? Yes.

Stirrups Standard? Yes.

Drawer Warmer? No.

Assist Arms? Yes.

Question #3 Does HCC have a minimum square footage requirement for the counseling rooms

Response #3 No

Question #4 Does HCC have a preference make/model for the vaccine refrigerator?

Response #4 Must be VFC Grade

Question #5 Does the restroom require a specimen pass-thru to the lab area?

Response #5 Yes

Question #6 Does the restroom require a toilet only or should a sink be included?

Response #6 Both

Question #7 Does HCC have a preference for location(s) of the desired privacy curtains?

Response #7 No

Question #8 Does HCC have a preference make/model for the medical-grade refrigerators? Or the required cu. ft?

Response #8 No. Compact preferred.

Question #9 Does HCC have a preference make/model for the medical-grade freezer? Or the required temperature? and the required cu. ft?

Response #9 No. Compact preferred.

Question #10 Will HCC be doing telehealth on via the desired communication package? If so, please specify the telehealth equipment and desired capabilities.

Response #10 Telehealth provided through HCC's Electronic Health Record.
eClinicalWorks/HEALOW and ZOOM.

Question #11 How many CAT 6 data outlets shall be included in the bid? What are the preferred locations?

Response #11 One in each area. Two in waiting area.

Question #12 Will HCC accept a TV larger than 24"?

Response #12 Yes

Question #13 What is the required location of the TV?

Response #13 Counseling and Exam Rooms

Question #14 What type of mobile cart is desired by HCC? Does HCC have a preference make/model of the mobile cart?

Response #14 No preference.

Question #15 Does HCC have a preference make/model for "vital signs equipment"?

Response #15 No preference.

Question #16 page 10 of 14 - Will a 33ft 50amp powered shore reel be acceptable?

Response #16 Yes